



USER INFORMATION

CE 2777

These products are classed as Category III Personal Protective Equipment (PPE) by the European PPE REGULATION 2016/425 and have been shown to comply with this Regulation through the Harmonised European Standard(s): EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018, EN ISO 374-5:2016

Product reference: X5000
Sizes available: M, L, XL, XXL
Colour: Blue

Intended Use:

Performance and limitation of use –This product has been tested and achieved the following performance levels:

Classification:

EN ISO 374-1:2016+A1:2018 /Type B	Level	EN ISO 374-4:2019 Degradation%
40% Sodium Hydroxide (K)	5	10
30% Hydrogen Peroxide (P)	2	36.5
37% Formaldehyde (T)	5	17.5
EN ISO 374-5:2016		
Protection against Bacteria and Fungi	Pass	
Protection against Viruses	Pass	

EN ISO 374-1:2016+A1:2018 /Type B



KPT



Virus

EN ISO 374-1:2016+A1:2018 Permeation levels are based on breakthrough times as follows:

Permeation performance level	1	2	3	4	5	6
Measured breakthrough time (min)	> 10	> 30	> 60	> 120	> 240	> 480

EN ISO 374-4:2019 Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.

EN ISO 374-5:2016 The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen.

This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals.

“The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal to or over 400 mm - where the cuff is tested also) and relates only to the chemical tested. It can be different if the chemical is used in a mixture.”

“It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on temperature, abrasion and degradation.”

“When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves”

"Before usage, inspect the gloves for any defect or imperfections."

Storage and transport: When not in use, store the product in a well-ventilated area away from extremes of temperature

Glove performance quoted is based on laboratory data and may not reflect the actual duration of protection in the workplace due to other factors influencing the performance such as temperature, abrasion, degradation etc.)

The glove does not contain any substances that are known to cause allergies.

The Gloves have no mechanical protection offered.

For single use only, do not littering.

Check for damage before use, do not use damaged gloves

Donning:

1. Remove all hand and wrist jewelry, and wash the hands before donning.
2. Place the gloves on the prepared work surface.
3. The user puts a glove on his/her dominant hand by grabbing it with the other hand, remembering to only touch the inside of the gloves, and slipping it over the dominant hand until it reaches final level.
4. The wearer uses the gloved dominant hand to slip the other glove onto the non-dominant hand.
5. Once both gloves are on, the users can touch the outside of the gloves to ensure a proper fit

Doffing:

1. Using the dominant hand, users start by grabbing the outside of the glove on the non-dominant hand on the palm side near the cuff.
2. Pull the glove off the non-dominant hand and place it in the gloved hand, balling it up.
3. Slip two fingers under the cuff of the other hand glove and carefully peel it off the hand without touching the wrist, turning the remaining glove inside out as it is removed and in turn encasing the first glove.
4. The gloves can be disposed.

Notified Body responsible for certification and ongoing conformity:

SATRA Technology Europe Ltd
Bracetown Business Park
Clonee, Dublin
D15 YN2P, Ireland (NB2777)

Product manufactured by: XINGYU MEDICAL TECH.CO.LTD

No. 2189 Yaoqian Road Gaomi Economic Development
Zone, Weifang City, Shandong Province, China



05.2024

XY202305LD

04.2027